


Determination of nutrients in seawater via NuLab Nutrient Analyser

April – June 2018

**Helmholtz-Zentrum Geesthacht
Institute for Coastal Research**

The Green Eyes NuLAB

Measurements Principles



Controller Unit

Reagents and OBS

Detector Units

Cd-Pipe

8-Port Valve

Waste

Syringe Pump

Standards

DI-Water


The Green Eyes NuLAB

Conducted Test



The Green Eyes NuLAB

Conducted Test – Laboratory Tests



	I	III	IV
Target value	100.24	26.11	1.80
Measured Mean	92.81	24.80	1.74
Deviation [$\mu\text{mol/l}$]	2.36	0.97	0.35
Deviation Error [%]	2.55	3.92	20.25
Error of Target [$\mu\text{mol/l}$]	7.43	1.31	0.06
Error of Target [%]	7.41	5.02	3.53

NOx Errors

	I	III	IV
Target value	5.21	2.78	0.64
Measured Mean	4.80	1.60	0.64
Deviation	0.24	0.42	0.06
Deviation Error [%]	4.94	7.79	5.30
Error of Target [$\mu\text{mol/l}$]	0.41	0.58	0.31
Error of Target [%]	7.94	8.62	47.32

NO2 Errors


	I	III	IV
Target value	5.04	2.10	0.50
Measured Mean	5.09	1.84	3.76
Deviation [$\mu\text{mol/l}$]	0.03	0.073	1.76
Deviation Error [%]	5.94	3.98	49.52
Error of Target [$\mu\text{mol/l}$]	0.05	0.27	3.76
Error of Target [%]	0.92	12.71	612.02

PO4 Errors


Parameter	Low Limit [$\mu\text{mol/l}$]	High Limit [$\mu\text{mol/l}$]
NO _x	0.8	200
NO ₂	0.6	150
PO ₄	0.2	25

The Green Eyes NuLAB

Conducted Test – Field Station Cuxhaven



Stability Test:



NO_x stability

1 calibration, 10 measurements

	NO ₂	NO _x	PO ₄
Target value	3.48	75.38	3.79
Mean	3.46	81.77	3.37
Deviation	0.09	2.45	0.19
% Error	2.67	3.00	5.53
Error of Target	0.02	-6.39	0.42
Error of Target %	0.71	-8.48	11.03

Overview 1 calibration, 1 measurement

	I	III	IV
Target value	100.24	26.11	1.80
Measured Mean	92.81	24.80	1.74
Deviation [$\mu\text{mol/l}$]	2.36	0.97	0.35
Deviation Error [%]	2.55	3.92	20.25
Error of Target [$\mu\text{mol/l}$]	7.43	1.31	0.06
Error of Target [%]	7.41	5.02	3.53

NOx Errors

	I	III	IV
Target value	5.21	2.18	0.65
Measured Mean	4.80	1.60	0.34
Deviation	0.24	0.12	0.02
Deviation Error [%]	4.94	7.59	5.30
Error of Target [$\mu\text{mol/l}$]	0.41	0.58	0.31
Error of Target [%]	7.94	26.62	47.32

NO2 Errors

	I	III	IV
Target value	5.04	2.10	0.50
Measured Mean	5.09	1.84	3.56
Deviation [$\mu\text{mol/l}$]	0.03	0.073	1.76
Deviation Error [%]	5.94	3.98	49.52
Error of Target [$\mu\text{mol/l}$]	0.05	0.27	3.06
Error of Target [%]	0.92	12.71	612.02

PO4 Errors

Overview 1 calibration, 10 measurements

	NO ₂	NO _x	PO ₄
Target value	3.48	75.38	3.79
Mean	3.46	81.77	3.37
Deviation	0.09	2.45	0.19
% Error	2.67	3.00	5.53
Error of Target	0.02	-6.39	0.42
Error of Target %	0.71	-8.48	11.03

- ✓ Good Precision
- ✓ Stable and reproducible Measurements
- ✓ Compact size (7 kg)
- ✓ Low Weight
- ✓ Separated Channels
- ✓ Macros can be adapted to needs and preferences
- ✓ Low Power Consumption
- ✓ Deployment and calibration remotely possible
- ✗ Carry-Over-Effect between sample and OBS
- ✗ No temperature normalisation on the NOx Channel
- ✗ A separated enclosure is necessary
- ✗ No washing cycle
- ✗ NOx reduction via Cadmium pipe
- ✗ Measurements could be faster
- ✗ Measuring the OBS without calibrating is missing as an option

Thank you for your
attention!